


=====
Acq. Operator : Seq. Line : 23
Acq. Instrument : Instrument 1 Location : Vial 19
Injection Date : 4/9/2018 10:32:50 PM Inj : 1
Inj Volume : 3.0 µl
Acq. Method : D:
Last changed : 3/24/2017 8:58:09 AM by 1
Analysis Method : D:\CHEM32\1\1.M
Last changed : 4/10/2018 9:01:40 AM
(modified after loading)
Sample Info : [A = ACN, B = 0.1% H3PO4/H2O] 0% A/B 2 min, 0-95% A/B 5 min, 95% A/B 3 min;
Athena C18 (3 µm, 2.1 mm * 100 mm); Flow: 0.45 ml/min, T = 25

Additional Info : Peak(s) manually integrated



=====
Area Percent Report
=====

Sorted By : Signal
Multiplier : 1.0000
Dilution : 1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: DAD1 B, Sig=280,4 Ref=400,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	4.517	BV	0.0107	2.92357	4.15280	0.0476
2	4.550	VV	0.0498	15.38732	4.29028	0.2503
3	4.644	VB	0.0116	3.33588	4.50940	0.0543
4	4.705	BB	0.0116	3.29026	4.46907	0.0535
5	6.827	BB	0.0417	6123.05029	2250.58081	99.5944

Totals : 6147.98733 2268.00237

=====
*** End of Report ***