


=====
Acq. Operator : Seq. Line : 4
Acq. Instrument : Instrument 1 Location : Vial 44
Injection Date : 5/21/2018 2:53:47 PM Inj : 1
Inj Volume : 3.0 µl
Acq. Method : D:
Last changed : 3/24/2017 8:58:09 AM by 1
Analysis Method : D:\CHEM32\1\1.M
Last changed : 5/21/2018 3:09:47 PM
(modified after loading)
Sample Info : [A = ACN, B = 0.1% H3PO4/H2O] 0% A/B 2 min, 0-95% A/B 5 min, 95% A/B 3 min;
Athena C18 (3 µm, 2.1 mm * 100 mm); Flow: 0.45 ml/min, T = 25

Additional Info : Peak(s) manually integrated



=====
Area Percent Report
=====

Sorted By : Signal
Multiplier : 1.0000
Dilution : 1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: DAD1 B, Sig=280,4 Ref=400,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	6.219	BB	0.0454	77.41339	26.22625	1.0891
2	6.694	BB	0.0555	31.73822	7.79134	0.4465
3	6.882	BB	0.0724	6999.01025	1557.19189	98.4644
Totals :				7108.16186	1591.20949	

=====
*** End of Report ***