


=====
Acq. Operator : Seq. Line : 3
Acq. Instrument : Instrument 1 Location : Vial 6
Injection Date : 7/24/2018 1:54:43 PM Inj : 1
Inj Volume : 3.0 µl
Acq. Method : D:
Last changed : 6/6/2018 1:07:35 PM
Analysis Method : D:\CHEM32\1\1.M
Last changed : 7/24/2018 2:27:00 PM
(modified after loading)
Sample Info : [A = ACN, B = 0.1% H3PO4/H2O] 0% A/B 2 min, 0-95% A/B 5 min, 95% A/B 3 min;
Athena C18 (3 µm, 2.1 mm * 100 mm); Flow: 0.45 ml/min, T = 25

Additional Info : Peak(s) manually integrated



=====
Area Percent Report
=====

Sorted By : Signal
Multiplier : 1.0000
Dilution : 1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: DAD1 C, Sig=210,4 Ref=400,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	6.789	BV	0.1009	12.50920	1.46854	0.5553
2	7.274	BV	0.0546	10.01097	2.33347	0.4444
3	7.518	VV	0.0545	5.69155	1.26515	0.2526
4	7.762	VB	0.0938	2216.12354	365.07480	98.3727
5	8.713	BV	0.0705	8.44829	1.43830	0.3750

Totals : 2252.78354 371.58026

=====
*** End of Report ***