


=====
Acq. Operator : Seq. Line : 1
Acq. Instrument : Instrument 1 Location : Vial 39
Injection Date : 9/27/2018 5:31:23 PM Inj : 1
Inj Volume : 3.0 µl
Different Inj Volume from Sequence ! Actual Inj Volume : 1.0 µl
Acq. Method : D:
Last changed : 8/9/2018 5:08:41 PM
Analysis Method : D:\CHEM32\1\1.M
Last changed : 9/27/2018 5:12:32 PM
(modified after loading)
Sample Info : [A = ACN, B = 0.1% H3PO4/H2O] 10-95% A/B 6 min, 95% A/B 4 min; Athena C18 (3
um, 2.1 mm * 100 mm); Flow: 0.45 ml/min, T = 40

Additional Info : Peak(s) manually integrated



=====
Area Percent Report
=====

Sorted By : Signal
Multiplier : 1.0000
Dilution : 1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: DAD1 B, Sig=280,4 Ref=400,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	6.758	BV	0.0382	3.10278	1.04408	0.0799
2	7.035	BB	0.0409	3.86656	1.31226	0.0996
3	7.413	BV	0.0348	3.34656	1.30934	0.0862
4	7.528	VV	0.0595	28.13039	6.97938	0.7245
5	7.750	VB	0.0812	3844.51343	653.08868	99.0099

Totals : 3882.95972 663.73374

=====
*** End of Report ***