



File D:\DATA\

Injection Date : 28 Apr 17 12:59 pm +0800 Tgt Mass(EZX) :
Sample Name : Location :
Acq. Operator : Inj : 1
Spec. Reported : MS Integration Inj Volume : 1 ul
Acq. Method : D:\METHODS\1-POS-3MIN.M
Analysis Method : D:\METHODS\1-POS-3MIN.M
Sample Info : Easy-Access Method: '1-POS-3MIN'
Method Info : Mobile Phase: A: water(10mMNH4HCO3) B: ACN
Gradient: 5%B increase to 95%B within1.5min,
95%B for 1.5min,back to 5%B within 0.01min.
Flow Rate: 1.8 mL/min
Column: XBridge, 3.5um, 50*4.6mm
Oven Temperature: 50 C


*DAD1 A, Sig=214,4 Ref=off



*DAD1 B, Sig=254,4 Ref=off



MSD1 TIC, MS File ES-API, Pos, Scan, Frag: 70



Integration Results for DAD1 A, Sig=214,4 Ref=off

RetTim	Width	Area	Height	Area%
1.75	0.02	4267.37	3303.55	100.00

Integration Results for DAD1 B, Sig=254,4 Ref=off

RetTim	Width	Area	Height	Area%
1.75	0.02	4452.13	3542.95	98.74
1.79	0.02	56.59	52.38	1.26

Ret. Time: 1.54 <<<< POSITIVE SPECTRA >>>>

ES-API Positive <<< Below Cutoff >>>

Ret. Time: 1.75

