


File D:\DATA\414
 Acq. Method : D:\METHODS\1-POS-3MIN.M
 Analysis Method : D:\METHODS\1-POS-3MIN.M
 Sample Info : *** No target masses specified ***
 Method Info : Mobile Phase: A: water (0.01%TFA) B: ACN (0.01%TFA)
 Gradient: 5%-95% B in 1.2min
 Flow Rate: 2.2 ml/min
 Column: Poroshell 120 EC-C18, 4.6*30mm, 2.7um
 Oven Temperature: 50 C
 Mass Range: 110-800



Integration Results for ADC1 A, ELSD

RetTim	Width	Area	Height	Area%
1.46	0.03	153.09	79.88	100.00

Integration Results for DAD1 A, Sig=214.4 Ref=off

RetTim	Width	Area	Height	Area%
1.46	0.03	2455.65	1231.42	99.46
1.72	0.01	13.24	23.54	0.54


Integration Results for DAD1 B, Sig=254.4 Ref=off

RetTim	Width	Area	Height	Area%
1.46	0.03	1303.61	683.59	100.00

LC/MS Report

Ret. Time: 1.48

<<<< POSITIVE SPECTRA >>>>



Ret. Time: 2.36

